

BLOOD GROUPING REAGENT

Anti-M

ALBAClone®

(Murine Monoclonal IgG)

For Tube Technique

REF Z171U

- **FOR IN VITRO DIAGNOSTIC USE**
- Meets FDA potency requirements
- Discard if turbid
- Preservative: 0.1% (w/v) sodium azide

CAUTIONS: THE ABSENCE OF ALL VIRUSES HAS NOT BEEN DETERMINED. THIS PRODUCT HAS COMPONENTS (DROPPER BULBS) CONTAINING DRY NATURAL RUBBER.

INTERPRETATION OF LABELING SYMBOLS



Batch code



Use by (YYYY-MM-DD)



Storage temperature limitation (2–8 °C)



In vitro diagnostic medical device



www.quantedia.com



Manufacturer



Product Code

INTENDED USE

This Anti-M reagent is for the *in vitro* detection and identification of the human M blood group antigen by direct agglutination.

SUMMARY AND EXPLANATION

The MN status of red blood cells is defined by the amino acid sequence of the major red cell sialoglycoprotein, glycophorin

A. Anti-M and anti-N react with their respective antigens on glycoprotein A, causing agglutination of the red blood cells and classifying these cells into three distinct phenotypes: M+N-, M+N+ and M-N+. Additionally, irrespective of the MN status of their major glycoprotein, almost all human red blood cells carry the 'N'-antigen on a minor red blood cell sialoglycoprotein, glycophorin B.

PRINCIPLE OF THE TEST

When used by the recommended technique, this reagent will cause the agglutination (clumping) of red blood cells carrying the M antigen. Lack of agglutination demonstrates the absence of the M antigen.

REAGENT DESCRIPTION

The main component of this reagent is derived from the *in vitro* culture of the IgG secreting mouse hybridoma:

Product Name	Product Code	Cell Line
Anti-M	Z171U	LMT

The formulation also contains bovine serum albumin, EPPS buffer and 0.1% (w/v) sodium azide.

NOTE: The volume delivered by the reagent bottle dropper is approximately 40 µL. Care should be taken to ensure that appropriate serum to cell ratios are maintained in all test systems.

STORAGE

The reagent should be stored at 2-8 °C.

WARNINGS AND PRECAUTIONS

For *in vitro* diagnostic use only

Products should be used by qualified personnel

Do not use beyond the expiration date

Do not use if turbid

Do not dilute

The format of the expiration date is expressed as YYYY-MM-DD (Year-Month-Day)

This reagent contains 0.1% (w/v) sodium azide. Sodium azide may be toxic if ingested and may react with lead and copper plumbing to form explosive compounds. If discarded into a sink, flush with a large volume of water to prevent azide buildup.

This reagent is of animal origin, therefore care must be taken during use and disposal as there is a potential infection risk.

CAUTION: SOURCE MATERIAL FROM WHICH THIS PRODUCT WAS DERIVED, WAS FOUND NEGATIVE FOR INFECTIOUS AGENTS WHEN TESTED IN ACCORDANCE WITH CURRENT FDA REQUIRED TESTS.

The bovine material which was used has been collected in a USDA approved facility.

Contains material of murine origin; therefore, handle appropriately as the absence of murine viruses has not been determined.

Monoclonal antibodies exhibit a high degree of potency, avidity and specificity. When using such antibodies, great care should be taken to avoid cross contamination.

This product has components (dropper bulbs) containing dry natural rubber.

SPECIMEN COLLECTION AND PREPARATION

Specimens should be collected by a standard collection technique. The specimen should be tested as soon as possible after collection. If testing is delayed, the specimen should be stored at refrigerated temperatures.

Clotted samples, or those collected in EDTA, should be tested within fourteen days from collection. Donor blood stored in citrate anticoagulant may be tested until the expiry date of the donation.

Special care should be taken if hemolyzed samples must be tested. Grossly icteric or contaminated blood specimens should not be used.

MATERIALS

Material provided

- ALBAClone® Anti-M

Materials required but not provided

- Unbuffered isotonic saline
- Reagent red blood cells suitable for the control of Anti-M
- 10 x 75 mm or 12 x 75 mm glass test tubes
- Pipettes
- Centrifuge
- Timer
- Heating block/waterbath (optional)

PROCEDURE

NOTE: This reagent has been standardized for use by the technique described below and therefore its suitability for use by other techniques cannot be guaranteed. When a test is required to be incubated for a specific time period, a timer should be used. It is recommended to allow the reagent to reach 20-25 °C prior to use.

When using supplemental testing equipment (i.e. centrifuge), follow the procedures that are contained in the operator's manual provided by the device manufacturer.

Tube Technique - 5 Minute Incubation/Spin

All red blood cells to be tested with this reagent should be washed at least once and resuspended in unbuffered isotonic saline. This includes red blood cells used for quality control.

1. Prepare a 2-4% suspension of red blood cells in unbuffered isotonic saline solution, 9 g/L NaCl.
2. Add 1 drop of blood grouping reagent to a glass test tube.
3. Add 1 drop of red blood cell suspension. Steps 2 and 3 may be performed in either order.
4. Mix the contents of the test tube and incubate at 20-25 °C for 5 minutes.

- Centrifuge the test tube.
NOTE: Suggested centrifugation: 900-1000 g (approx. 3400 rpm) for 10 seconds or a time and speed appropriate for the centrifuge used that produces the strongest reaction of antibody with antigen-positive red blood cells, yet allows easy re-suspension of antigen-negative red blood cells.
- After centrifugation, gently shake the tube to dislodge the cell button from the bottom and immediately observe macroscopically for agglutination. Do not use any optical aid to examine the testis results.
- Record results.

STABILITY OF REACTION

Test results should be read, interpreted and recorded immediately after centrifugation. Delays may cause dissociation of antigen-antibody complexes resulting in weak positive or false negative reactions.

INTERPRETATION OF RESULTS

Agglutination = positive test result
No agglutination = negative test result

QUALITY CONTROL

Quality control of reagents is essential and should be performed on each day of use and in accordance with local, state and federal regulations.

M+N+ red blood cells should be used as a positive control
M-N+ red blood cells should be used as a negative control

False positive test results are rarely seen with low-protein reagents. False positive agglutination may be due to a positive direct antiglobulin test (DAT), cold agglutinins, or abnormal serum proteins. If false positive results are suspected, or local regulations require, and a control test for spontaneous agglutination is desired, ALBACheck® - BGS Monoclonal Control (ZZ71U) or 6-10% albumin in saline may be substituted for the blood grouping reagent in the testing procedure. A negative result would serve as an appropriate control. If the monoclonal control test gives a positive reaction, a valid interpretation of the results obtained in red blood cell testing cannot be made without further investigation.

LIMITATIONS

As this reagent reacts optimally at pH 8.5 and is extremely sensitive to pH, test red blood cells should be suspended in unbuffered medium. All red blood cells suspended in buffered medium e.g. Modified Alsever's solution, should be washed at least once and resuspended in unbuffered saline prior to use.

Incubation at temperatures above that recommended may result in weaker reactions.

Cells modified by proteolytic enzymes must not be used, as M antigens may be destroyed.

Excessive centrifugation can lead to difficulty in resuspending the cell button, while inadequate centrifugation may result in agglutinates that are easily dispersed.

The expression of certain red blood cell antigens may diminish in strength during storage, particularly in EDTA and clotted samples. Better results will be obtained with fresh samples.

False positive or false negative results can occur due to contamination of test materials, improper reaction temperature, improper storage of materials, omission of test reagents and certain disease states.

Suppressed or weak expression of blood group antigens may give rise to false negative reactions

Gently re-suspend tube tests before reading. Excessive agitation may disrupt weak agglutination and produce false negative results.

SPECIFIC PERFORMANCE CHARACTERISTICS

Prior to release, each lot of ALBACheck® Anti-M is tested using FDA recommended methods against a panel of antigen-positive and antigen-negative red blood cells to ensure suitable reactivity.

During comparator studies (data on file at Alba Bioscience Limited), blood samples were tested with ALBACheck® Anti-M (Monoclonal) as follows*:

Anti-M	Comparator Reagent		One-sided 95% Exact lower confidence limit
	Positive	Negative	
Positive	81	0	81
Negative	0	19	19
Total	81	19	100
Positive Percent Agreement*			100 96.37
Negative Percent Agreement*			100 85.41

* The data presented in this table was generated during field trials executed in support of the original US licensing of this reagent.

BIBLIOGRAPHY

- Roback JD, Grossman BJ, Harris T, et al. AABB Technical Manual, ed 18. AABB, 2014
- AABB Standards Program Committee: Standards for Blood Banks and Transfusion Services, ed 30. AABB, 2016
- Reid ME, Lomas-Francis C, Olsson ML: The Blood Group Antigen FactsBook, ed 3. Academic Press, 2012

DATE OF ISSUE

2020-04-07

US Distributor

Quotient
301 South State Street
S-204
Newtown
PA 18940
USA

Customer Service Tel: 1-888-284-1901
Product Technical Support Tel: 1-888-228-1990
Customer Service Fax: 1-888-694-5208
E-Mail: customer.service.US@quotientibd.com
Web: www.quotientibd.com/us



Alba Bioscience Limited
James Hamilton Way
Penicuik
EH26 0BF
UK
U.S. License 1807

© Alba Bioscience Limited 2020

Z171UPI/06

